Set de verificación de calibración para HI718

HI718-11

Descripción

El Set de verificación de calibración Yodo Checker® HC - **HI718-11** proporciona una solución simple para validar el Iodine Checker® HC - **HI718** que tiene un rango de 0.0 a 12.5 ppm (mg/l). Este conjunto de estándares de alta calidad se fabrica en nuestras instalaciones de última generación y se suministra con un certificado de análisis. El Certificado de análisis proporciona el número de lote, los valores de referencia y la fecha de vencimiento para la trazabilidad al certificar el Checker® HC.

Especificaciones

No Especifica

Accesorios

No Especifica

Cómo pedir

No Especifica

Ventajas

El HI 718-11 se suministra con Certificado de Análisis

- Código de producto
- Número de lote
- Fecha de caducidad

HANNA instruments

CERTIFICATE OF ANALYSIS

Product Name: CALIBRATION CHECKING SET FOR HI 711
Product Code: HI 711-11
Lot number: SC0372/14
Expiration date: [illegible]ber 2017

Standard cuvette	Lot number	Standard Value @ 25°C (ppm)	Passed
A, HI 711-11	2038	[illegible]	✓
B, HI 711-11	2039	[illegible] ± 0.05	✓

Hanna Instruments certifies these standards [illegible] stated tolerance limits as indicated. These should be used for [illegible]y contr[illegible] to validate instrument function on the specified Hanna in[illegible] only. [illegible]re standards in the storage box upright at room tempera[illegible]e. Do n[illegible] standard cuvettes.

Instructions for use:

1. Thoroughly wipe standard cuvette with HI731318 [illegible] free cloth and only handle using top of cuvette. The cuvette wall must be dry and free of smudges.
2. At the C.1 prompt place A, HI 711-11 standard cuvette into the h[illegible] of the Checker and press operations button.
3. At the C.2 prompt place B, HI 711-11 standard cuvette i[illegible] holder an[illegible] press the operations button.

Acceptable results must be between 0.95 and 1.05 ppm @ [illegible]°C.

File number: CERT711-11 _SC0372/14
QA manager: Bogdan Munteanu

SAMPLE

Video

No Especifica